


Sample Description:

Structure : DM-7
 Column : 4.6×250mm, Sinochrom ODS-BP-5
 Solvent A : 0.1% trifluoroacetic in 100% acetonitrile
 Solvent B : 0.1% trifluoroacetic in 100% water
 Gradient :
 A B
 0.01min 22% 78%
 25min 45% 55%
 25.1min 100% 0%
 30min STOP
 Flow rate : 1.0 mL/min
 Wavelength : 220nm

Cat. No.: RP10516
 Lot No.: PX00511503



Results

Peak No.	Peak ID	Ret Time	Height	Area	Conc.
1		13.833	1261.981	10934.792	0.2651
2		14.360	1903.726	14529.394	0.3522
3		14.697	4053.019	50187.188	1.2165
4		15.123	5341.241	11094.095	0.2689
5		15.298	299289.469	3993997.250	96.8146
6		15.868	2355.899	23465.943	0.5688
7		16.203	1340.248	10987.940	0.2663
8		16.993	622.557	6450.549	0.1564
9		17.703	441.814	3759.545	0.0911
Total			316609.953	4125406.696	100.0000