


Summary:

Structure : FQ-22  
 Column : 4.6×250mm, Kromasil C18-5  
 Solvent A : 0.1% trifluoroacetic in 100% acetonitrile  
 Solvent B : 0.1% trifluoroacetic in 100% water  
 Gradient :  
           A          B  
 0.01min 20%      80%  
 25min   45%      55%  
 25.1min 100%     0%  
 30min          STOP  
 Flow rate : 1.0 mL/min  
 Wavelength : 220nm

Cat. No. : RP10526  
 Lot No. : PX00105502



RESULTS

Peak No.	Peak ID	Ret Time	Height	Area	Conc
1		14.883	541.804	3839.845	0.2015
2		16.827	547.548	6887.100	0.3614
3		18.190	138306.172	1860075.875	97.5999
4		18.688	4479.546	35014.996	1.8373
TOTAL			143875.070	1905817.817	100.0000