


Sample Name : Bradykinin
 Cat. No. : RP11129

Pump A : 0.065% trifluoroacetic in 100% water
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile
 Total Flow : 1 ml / min
 Wavelength : 214 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.10	Pumps	Pump A B.Conc	95	
33.00	Pumps	Pump A B.Conc	95	
33.10	Pumps	Pump A B.Conc	5	
43.00	Pumps	Pump A B.Conc	5	
43.10	Controller	Stop		

<Detector A>
 Column : X Bridge 4.6 x 250mm



1 Det.A Ch1 / 214nm

Peak Table

Detector A Ch1 214nm

Peak#	Ret. Time	Area	Height	Area %
1	9.729	8990	1749	0.028
2	10.550	16829	3374	0.052
3	10.937	36668	7302	0.114
4	14.145	15014	3093	0.047
5	14.794	643941	102261	2.004
6	15.016	30965772	2270907	96.355
7	15.325	252168	51379	0.785
8	15.458	53837	10159	0.168
9	15.634	107442	20914	0.334
10	15.971	14801	3006	0.046
11	19.863	21668	4210	0.067
Total		32137129	2478354	100.000