


Summary:

Structure : SV-13
 Column : 4.6×250mm, Kromasil C18
 Solvent A : 0.1% trifluoroacetic in 100% acetonitrile
 Solvent B : 0.1% trifluoroacetic in 100% water
 Gradient :
 A B
 0.01min 20% 80%
 25min 45% 55%
 25.1min 100% 0%
 30min STOP

Cat. No.: RP11300
 Lot No.: PJ01125504

Flow rate : 1.0 mL/min
 Wavelength : 220nm



RESULTS

Peak No.	Peak ID	Ret Time	Height	Area	Conc
1		13.097	268.981	1706.351	0.1882
2		13.400	609.915	5028.306	0.5545
3		14.007	82870.070	868595.938	95.7811
4		14.478	986.261	8816.603	0.9722
5		14.743	1387.291	14147.504	1.5601
6		15.868	986.092	8560.950	0.9440
TOTAL			87108.610	906855.651	100.0000