


Sample Description:

Structure : VP-13
 Column : 4.6×250mm, Kromasil-C18-5
 Solvent A : 0.1% trifluoroacetic in 100% acetonitrile
 Solvent B : 0.1% trifluoroacetic in 100% water
 Gradient :
 A B
 0.01min 17% 83%
 25min 32% 68%
 25.1min 100% 0%
 30min STOP

Cat. No.: RP11641
 Lot No.: PY00213507

Flow rate : 1.0 mL/min
 Wavelength : 220nm



Results

Peak No.	Peak ID	Ret Time	Height	Area	Conc.
1		18.003	1426.780	23058.855	0.0994
2		18.662	27464.375	576709.250	2.4858
3		19.128	7370.884	79445.922	0.3424
4		19.620	30026.969	389551.875	1.6791
5		20.183	709414.125	22130948.000	95.3932
Total			775703.133	23199713.902	100.0000