

Sample Name : Binin(102-120), Strongylocentrotus Purpuratus
 Cat No. : RP19984


Pump A : 0.065% trifluoroacetic in 100% water
 Pump A : 0.05% trifluoroacetic in 100% acetonitrile
 TotalFlow:1 ml/min
 Wavelength:214 nm

Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
30.00	Pumps	Pump A B.Conc	100	
42.00	Pumps	Pump A B.Conc	100	
43.00	Pumps	Pump A B.Conc	5	
49.00	Pumps	Pump A B.Conc	5	
50.00	Controller	Stop		

<Detector A>

Column : SHIMADZU VP-ODS 4.6 x 150 mm

Chromatogram



1 Det.A Ch1 / 214nm

Peak Table

Detector A Ch1 214nm

Peak#	Ret. Time	Area	Height	Area %
1	11.974	48272	5318	0.339
2	12.841	69879	8137	0.491
3	13.471	21334	3013	0.150
4	14.160	12544	1154	0.088
5	15.815	18278	2510	0.129
6	16.665	14020788	2110890	98.607
7	17.058	27692	5501	0.195
Total		14218787	2136524	100.000