


Sample Description:
 Structure : YG-22
 Column : 4.6×250mm, Sinochrom ODS-BP-5
 Solvent A : 0.1% trifluoroacetic acid in 100% acetonitrile
 Solvent B : 0.1% trifluoroacetic acid in 100% water
 Gradient :
 A B
 0.01min 23% 77%
 25min 50% 50%
 25.1min 100% 0%
 30min STOP

RP10424
 PX00205509

Flow rate : 1.0mL/min
 Wavelength : 220nm



Results

Peak No.	Peak ID	Ret Time	Height	Area	Conc.
1		10.845	535.549	3671.795	0.0566
2		11.303	677.303	4607.090	0.0710
3		11.607	1256.042	7376.484	0.1137
4		12.257	544228.938	6459773.000	99.6001
5		14.417	1097.355	10278.818	0.1585
Total			547795.188	6485707.188	100.0000