


MS Spectrum



Sample Information Item
Date and Time : 2005-3-24 18:35:29
User : FJS
Sample : RP13717
Inj. Volume : 1.000
MW : 1181.42

Probe: ESI
Nebulizer Gas Flow: 1.5L/min
CDL: -20.0v
Block Temp: 250°C

Probe bias: +4.5kv
Detector: 1.5kv
T. Flow: 0.2ml/min
B. conc: 50%H2O/50%ACN

RP13717
PY00308508