


MS Spectrum



Sample Information Item : 2005-5-19 20:51:09
Date and Time : xie
User : RP10491
Sample : RP10491
Inj. Volume : 1.000
MW : 4011.47

Probe bias: +4.5kv
Detector: 1.5kv
T. Flow: 0.2ml/min
B. conc: 50%H2O/50%ACN

Probe: ESI
Nebulizer Gas Flow: 1.5L/min
CDL: -20.0v
CDL Temp: 250°C
Block Temp: 200°C